

CR / DR QC Contrast Detail Phantom Evaluation

Cassette Number:

Initial Technique Settings and Score of QC Phantom									
SID	kVp	mAs	mA	mA Time (s) Focal Spot					
	Circle Score			EI / S-Value					

Any subsequent score value under initial phantom score is a failing score.

 $For any \ failing \ scores, \ review \ technique \ settings \ and \ contact \ service \ for \ further \ troubleshooting.$

Date (mm/dd/yy)	# of Circles	EI / S-Value	Pass (x)	Fail (x)	Comments	Initials
_						



Date (mm/dd/yy)	# of Circles	EI / S-Value	Pass (x)	Fail (x)	Comments	Initials